



Bid Number/बोली क्रमांक (बिड संख्या):
GEM/2025/B/5912627
Dated/दिनांक : 05-02-2025

Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	20-02-2025 15:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	20-02-2025 15:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	180 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Agriculture And Farmers Welfare
Department Name/विभाग का नाम	Department Of Agricultural Research And Education (dare)
Organisation Name/संगठन का नाम	Indian Council Of Agricultural Research (icar)
Office Name/कार्यालय का नाम	Icar-directorate Of Groundnut Research
क्रैता ईमेल/Buyer Email	buyer192.icari.gj@gembuyer.in
Total Quantity/कुल मात्रा	1
Item Category/मद केटेगरी	Gas Chromatography (Q3)
MSE Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से एमएसई छूट	No
Startup Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से स्टार्टअप छूट	No
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Certificate (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Do you want to show documents uploaded by bidders to all bidders participated in bid?/	No
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	No
Type of Bid/बिड का प्रकार	Two Packet Bid
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	2 Days
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No

Bid Details/बिड विवरण

Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation
Arbitration Clause	No
Mediation Clause	No

EMD Detail/ईएमडी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
EMD Amount/ईएमडी राशि	100000

ePBG Detail/ईपीबीजी विवरण

Advisory Bank/एडवाइजरी बैंक	State Bank of India
ePBG Percentage(%) / ईपीबीजी प्रतिशत (%)	5.00
Duration of ePBG required (Months) / ईपीबीजी की अपेक्षित अवधि (महीने).	50

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy./जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कैटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई कैटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

Beneficiary/लाभार्थी :

DIRECTOR

Icar-directorate Of Groundnut Research, Department of Agricultural Research and Education (DARE), Indian Council of Agricultural Research (ICAR), Ministry of Agriculture and Farmers Welfare
(The Director ligr)

MII Purchase Preference/एमआईआई खरीद वरीयता

MII Purchase Preference/एमआईआई खरीद वरीयता	Yes
--	-----

MSE Purchase Preference/एमएसई खरीद वरीयता

MSE Purchase Preference/एमएसई खरीद वरीयता	Yes
---	-----

1. Preference to Make In India products (For bids < 200 Crore): Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

[OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

2. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

Gas Chromatography (1 pieces)

(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/कमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)

Technical Specifications/तकनीकी विशिष्टियाँ

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
GAS CHROMATOGRAPH (GC) CONFIGURATION	"GC Detectors inclusive in the scope of supply (Select Only those Detectors, which are part of supply)"	Flame Ionization Detector (FID), Thermal Conductivity Detector (TCD or Hot Wire Detector), Electron Capture Detector (ECD), Nitrogen-Phosphorus Detector (NPD) / Flame Thermionic Detector (FTD), Flame Photometric Detector (FPD), Photo Ionization Detector (PID), Pulsed Flame Photometric Detector (PFPD)
	"Number of FID in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	"Number of TCD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	"Number of ECD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	"Number of NPD/FTD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	"Number of FPD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	"Number of PID in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	"Number of PFPD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0, 1, 2
	Number of Inlets GC can Simultaneously Support	1, 2, 3
	Number of Detectors (Ports) GC can Simultaneously Support	1, 2, 3, 4
	Number of Capillary Columns provided in GC - inclusive in the scope of Supply	0, 1, 2
	Type of Capillary Column	Polar, Non-Polar, Semi-Polar, NA
	"GC Capillary Column Length (Hint:-Select '0' if Not Applicable)"	0, 30, 60, 100 (meter)
	"Number of Packed Columns provided in GC - inclusive in the scope of Supply (Hint:-Select '0' if Not Applicable)"	0, 1, 2

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	"GC Packed Column Length (Hint:-Select '0' if Not Applicable)"	0, 3, 6, 9, 12 (foot)
PERFORMANCE	"GC Detector Minimum Detection Level (MDL) "	< 1.4 pg C/s for FID, < 1.8 pg C/s for FID, < 3.0 pg C/s for FID < 400 pg/ml for TCD, < = 1 microgram/ml for TCD, < = 10 microgram/ml for TCD, < 4.5 fg /sec for ECD, < 5.5 fg /sec for ECD, <= 100 fg/sec for ECD, < 0.1pg N /sec for NPD/FTD, < 0.03 pg P /sec for NPD/FTD, <0.5 pg P/s for NPD/FTD, < 45 fg P /sec for FPD, < 100 fg P /sec for FPD, <2.5 pg S /sec for FPD, <2.5 pg S /sec for FPD, <5 pg S /sec for FPD, <10pg S/s for FPD
COLUMN OVEN (GAS CHROMATOGRAPH - GC)	Maximum Operating Temperature (GC Oven)	upto 400 deg. C, upto 450 deg. C (2) Or higher
	Number of Oven Ramps / Plateaus for Temperature programming	6/7 or more, 9/10 or more, 20/21 or more
	Temperature Ramp Rate - Maximum	120 Deg. C/minute or better, 50 Deg. C/minute or better
PNEUMATIC CONTROL	Number of Injection Ports	1, 2
	Maximum Injection Port heating Temperature	upto 400 deg. C (1), upto 450 deg. C (2) Or higher
	Pressure Range	upto100 psi (1), upto 120 psi (2) Or higher
	Pressure set point Resolution (\pm)(psi)	0.001, 0.01, 0.1 Or lower
	Type of Sample Injection Port	Split, Splitless, split / Splitless, Purged packed Inlet
	Split Ratio	6000:1, 7000:1, 7500:1, NA
AUTO INJECTOR / AUTO SAMPLER	Compatible Liquid Auto Sampler - inclusive in the scope of supply	Yes, No
	Capacity of the Liquid Auto Sampler (Number of Vials)	0, 15, 100, 150 Or higher
	Number of Injection Ports in the Liquid Auto Sampler	0, 1, 2
	Number of Channels in the Liquid Auto Sampler	0, 2
	Minimum Volume of the injected liquid sample	0 - 1 (microliter)
	Maximum Volume of the injected liquid sample	0 - 100 (microliter)
	"Liquid Auto Sampler Features (Select applicable features Only)"	Fully automated Liquid Auto Sampler, Injects with variable speed & varying syringe sizes, Operates through Software (Software inclusive in the scope of supply), NA

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
HEAD SPACE SAMPLER	Compatible Head Space Sampler- inclusive in the scope of supply	Yes, No
	Capacity of the Head Space Sampler Carousel (number of Vials)	12 or better, NA
	Head Space Sampler Temperature Range	Upto 200 deg. C or better, NA
PYROLYZER SYSTEM	Compatible Pyrolyzer System for pyrolysis of solid and liquid samples- inclusive in the scope of supply	Yes, No
SALIENT FEATURES OF GAS CHROMATOGRAPH (GC)	"Salient Features of GC (Select applicable features Only)"	Automatic leak testing/detection of Pneumatics, Direct setting of split flow rates and ratios, Flow sensor for control & storage of split ratio, Split and Septum Purge gases, Able to direct inject onto packed and wide-bore capillary columns, Oven power turns-off automatically when the lid/door is opened, Over Heat Protection, Display Unit with built-in Key Board, Controlled by External PC System, NA
SOFTWARE	"Software Features (Select Applicable Features only)"	Compatible with Windows 7 or higher version operating system, Seamless integration & control of all the GC parameters, Interface with Multiple GCs through the comprehensive real-time monitoring and diagnostic software, Real time notification via advisories and indicators, Counters, Chromatographic attributes, Instrument diagnostics, Access to maintenance information, Access to maintenance, run and event logs, Retention Time - Locking module for analyzing target compounds in complex matrices
GENERIC	Display	Through PC (PC is not part of supply), Built-in Display
	Connectivity interface	USB, WIFI, ETHERNET, Bluetooth, RS 232, RJ-45
	Provision to split the sample from single injection to two different detectors and get results for two detectors simultaneously	Yes, No
ACCESSORIES	Gas Supply inclusive in the scope of supply	Nitrogen, Hydrogen, Argon, Helium, Zero Air, Not offered
	Capacity of Gas cylinder	47 liters water capacity, Not offered
	Number of Cylinders of each type supplied - inclusive in the scope of supply	0, 1, 2
	Gas cylinders certified by PESO	Yes, No for Gas Cylinder Not Offered
	Regulator - inclusive in the scope of supply	Dual Stage Regulator, NA

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	"Other Accessories inclusive in the scope of supply (Select only those Accessories which are part of supply)"	Gas purification panel for all Gases with moisture trap, Hydrocarbon trap & Oxygen trap, Syringe for manual injection x 05 nos, Auto Sampler Vials x 1000 vials with screw cap, Graphite ferrules x 10 nos, Vessel ferrules x 10 nos, Septa for injector x 100 nos, Inlet liner for splitless x 4 nos, Inlet liner for split x 4 nos, O-rings x 10 nos. of each type, Column Cutter x 2 nos, Gas tube Cutter x 2 nos, Tool Kit x 1 no., User / Technical / Maintenance Manual, Standard Samples - NIST / IPL Traceable, NA
ENVIRONMENTAL	Minimum Operating Temperature	0 - 10 (degree Celsius)
	Maximum Operating Temperature	30 - 65 (degree Celsius)
	Operating Humidity (RH) (%) at 40 degree C	Up to 95 (percent)
INSTALLATION & COMMISSIONING / WARRANTY / AVAILABILITY OF SPARES / TRAINING	Warranty on equipment	1, 2, 3, 5 (year)
	Number of days onsite training provided	3, 5, 7, 10 (day)
CERTIFICATION / TEST REPORT	"Availability of UL/CE Certificate as per IEC 61010- 1 latest - Safety requirements for electrical equipment for measurement, control, and laboratory use General requirements"	Yes, No
	Conformance to EMC/EMI as per IEC 61326-1 latest - Electrical equipment for measurement, control and laboratory use	Yes, No
	Availability of Test Reports from Central Govt / NABL approved / ILAC accredited Lab to prove conformity to the specification	Yes, No

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Dilip Kumar	362001,PB No-5, Ivnagar Road	1	15

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 % of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 25% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$, subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

2. OEM

IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

3. Warranty

Successful bidder will have to ensure that adequate number of dedicated technical service personals / engineers are designated / deployed for attending to the Service Request in a time bound manner and for ensuring Timely Servicing / rectification of defects during warranty period, as per Service level agreement indicated in the relevant clause of the bid.

4. Generic

Without prejudice to Buyer's right to price adjustment by way of discount or any other right or remedy available to Buyer, Buyer may terminate the Contract or any part thereof by a written notice to the Seller, if:

- The Seller fails to comply with any material term of the Contract.
- The Seller informs Buyer of its inability to deliver the Material(s) or any part thereof within the stipulated Delivery Period or such inability otherwise becomes apparent.
- The Seller fails to deliver the Material(s) or any part thereof within the stipulated Delivery Period and/or to replace/rectify any rejected or defective Material(s) promptly.
- The Seller becomes bankrupt or goes into liquidation.
- The Seller makes a general assignment for the benefit of creditors.
- A receiver is appointed for any substantial property owned by the Seller.
- The Seller has misrepresented to Buyer, acting on which misrepresentation Buyer has placed the Purchase Order on the Seller.

5. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

6. Forms of EMD and PBG

Successful Bidder can submit the Performance Security in the form of Fixed Deposit Receipt also (besides PBG which is allowed as per GeM GTC). FDR should be made out or pledged in the name of

The Director IIGR-Junagadh

A/C (Name of the Seller). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of Security Deposit, the FDR will be released in favour of bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Successful Bidder has to upload scanned copy of the FDR document in place of PBG and has to ensure delivery of hard copy of Original FDR to the Buyer within 15 days of award of contract.

7. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

The participants must upload the documents as per the specifications mentioned in the Bid ATC

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976,

The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

---Thank You/धन्यवाद---